

# 2020 ANNUAL SUMMARY REPORT


Newborn Screening  
Quality Assurance  
Program



**Centers for Disease  
Control and Prevention**  
National Center for  
Environmental Health

# Newborn Screening Quality Assurance Program 2020 Annual Summary Report, Volume 38

U.S. Department of Health and Human Services  
Centers for Disease Control and Prevention  
National Center for Environmental Health  
**Division of Laboratory Sciences**



**Note for accessibility: Explanations Figures 2–37 (bias plots) are located in [Appendix for Accessibility Descriptions, page 43](#).**

## Contents

Acronym Glossary.....	ii
Introduction.....	1
New Activities.....	2
About NSQAP.....	3
Countries Shown on World Map that Participated in NSQAP During 2020 .....	4
Filter Paper.....	6
Proficiency Testing.....	7
The Proficiency Testing Analytes .....	7
Proficiency Testing Materials and Methods.....	8
Proficiency Testing Data Handling.....	8
Proficiency Testing Errors.....	8
Proficiency Testing Cutoff Values.....	14
2020 Bias Plots.....	24
Proficiency Testing Bias Plots.....	24
Appendix for Accessibility Descriptions.....	43
References.....	43
Acknowledgments.....	44

# Acronym Glossary

Notation	Description
<b>17OHP</b>	17 $\alpha$ -hydroxyprogesterone
<b>A2LA</b>	American Association for Laboratory Accreditation
<b>ALD</b>	adrenoleukodystrophy
<b>BIOT</b>	biotinidase
<b>BMSL</b>	Biochemical Mass Spectrometry Laboratory
<b>CAH</b>	second-tier congenital adrenal hyperplasia
<b>CDC</b>	Centers for Disease Control and Prevention
<b>CFDNA</b>	cystic fibrosis DNA
<b>CLSI</b>	Clinical Laboratory Standards Institute
<b>DBS</b>	dried blood spot
<b>EV</b>	expected value
<b>FDA</b>	Food and Drug Administration
<b>G6PD</b>	glucose-6-phosphate dehydrogenase
<b>GALT</b>	galactose-1-phosphate uridylyltransferase
<b>Hb</b>	sickle cell and other hemoglobinopathies
<b>HIV</b>	anti-human immunodeficiency virus-1 Antibody
<b>HORM</b>	hormone + total galactose
<b>IEC</b>	International Electrotechnical Commission
<b>IRT</b>	immunoreactive trypsinogen
<b>ISO</b>	International Organization for Standardization
<b>LC</b>	liquid chromatography
<b>LSD</b>	lysosomal storage disorder
<b>MAP</b>	Molecular Assessment Program
<b>MQIP</b>	Molecular Quality Improvement Program
<b>MS/MS</b>	tandem mass spectrometry
<b>MSMS1</b>	tandem MS 1
<b>NSMBB</b>	Newborn Screening and Molecular Biology Branch
<b>NSQAP</b>	Newborn Screening Quality Assurance Program
<b>PT</b>	proficiency testing
<b>QC</b>	quality control
<b>RBC</b>	red blood cells
<b>RUSP</b>	Recommended Uniform Screening Panel
<b>SMA</b>	Spinal Muscular Atrophy
<b>SMN1</b>	survival motor neuron 1
<b>T4</b>	thyroxine
<b>TGAL</b>	total galactose
<b>TOXO</b>	anti-Toxoplasma Antibody
<b>TREC</b>	T-cell receptor excision circle
<b>TSH</b>	thyroid stimulating hormone

# Newborn screening is one of the most successful preventative health programs in the United States.



## Introduction

Newborn screening is one of the most successful preventative health programs in the United States. Healthcare professionals collect dried blood spot (DBS) specimens from more than 98% of all US newborns shortly after birth. State and public health laboratories or associated laboratories screen these DBS specimens for certain genetic, metabolic, and endocrine disorders. The Centers for Disease Control and Prevention's (CDC) Newborn Screening Quality Assurance Program (NSQAP) helps newborn screening laboratories with these testing processes.

NSQAP produces certified DBS materials for proficiency testing (PT) and quality control (QC) analysis, works to improve the quality and scope of laboratory services, and provides consultation to laboratories. State-operated and private newborn screening laboratories process thousands of DBS specimens daily. NSQAP helps newborn screening laboratories ensure that testing

accurately detects disorders, does not delay diagnoses, minimizes false-positive reports, and sustains high-quality performance.

CDC's Newborn Screening and Molecular Biology Branch (NSMBB) has been granted International Organization for Standardization (ISO)/International Electrotechnical Commission (IEC) 17043 accreditation by the American Association for Laboratory Accreditation (A2LA). Accreditation was achieved after a thorough review of NSMBB's quality management system and ability to develop and administer specific PT protocols. The branch's NSQAP web-based PT programs are included in the A2LA Scope of Accreditation. The accreditation does not include testing for glucose-6-phosphate dehydrogenase (G6PD) and NSQAP disease specific PT programs. Consult [A2LA Certificate#4190.01](#) for a list of accredited NSMBB PT programs.

# New Activities

- January – NSQAP launched the Participant Portal (<https://nbs.dynamics365portals.us/>) which allowed for direct entry of results by participants in the areas of:
  - PT – Participants entered results into both the new portal and legacy site for the following programs: amino acid PT, acylcarnitine PT, biotinidase (BIOT) PT, glucose-6-phosphate dehydrogenase (G6PD) PT, hormone + total galactose (HORM) PT, galactose-1-phosphate uridylyltransferase (GALT) PT, immunoreactive trypsinogen (IRT) PT. Comparative analysis showed negligible differences between the modes of data entry.
    - Interactive Power BI reports for biochemical PT statistical reports were introduced. These reports allowed participants to view informational statistical data by specimen, by analyte, or by location of labs (US domestic or international).
    - Individual laboratory evaluations were enhanced. Specimens were evaluated as “Acceptable” or “Unacceptable”. If the participant’s reported clinical assessment differed from the NSQAP expected clinical assessment, the evaluation for the specimen was marked “Unacceptable”. It was the responsibility of the laboratory to categorize unacceptable results according to their protocols.
    - Laboratory evaluations and reports were distributed directly to the NSQAP portal. Participants with access to the NSQAP portal viewed and/or downloaded reports and evaluations.
  - QC - Pilot testing of the NSQAP Participant Portal was introduced to a subset of domestic participants to validate its functionality for QC data entry of the following programs: 17  $\alpha$ -hydroxyprogesterone + total galactose (17OHP + TGAL) QC, lysosomal storage disorder (LSD) QC, tandem MS 1 (MSMS1) QC, thyroxine (T4) QC, thyroid stimulating hormone (TSH)QC, along with second-tier QC programs for congenital adrenal hyperplasia (CAH), maple syrup urine disease and phenylketonuria (MSUD-PKU), and methylmalonic/propionic acidemia and homocystinuria (MMA-tHCY).
- March - NSQAP and participating laboratories experienced unprecedented challenges due to the COVID-19 pandemic. CDC safety precautions were introduced to slow the spread of the disease and resulted in staffing and material production limitations, necessitating an adjustment to the routine material shipment schedule. Despite these circumstances, NSQAP provided two PT events and two QC events to all participants during the second half of 2020.
- July - A new PT program for Spinal Muscular Atrophy (SMA) was piloted. This program used DBS to determine the presence or absence of survival motor neuron 1 (SMN1) exon 7. The pilot program was a success and the SMA PT program will be available for all participants in 2021. For more information or to request participation in this program, email [NSQAPDMT@cdc.gov](mailto:NSQAPDMT@cdc.gov).
- August through December – Official QC program data collection via the NSQAP Participant Portal was conducted for all participants. For this data set, statistical enhancements were developed to better access method peer group performance and will be available when the 2020 QC Summary Report for Set 2 is published.

## About NSQAP

For more than 40 years, NSQAP and its cosponsor, the Association of Public Health Laboratories, have researched the development of DBS quality assurance materials for newborn screening tests and have assisted laboratories with DBS-related testing issues. NSQAP primarily supports US newborn screening laboratories; however, private and international laboratories can enroll in the program. Participation is voluntary. NSQAP provides quality assurance services for the core (primary) and secondary conditions listed in the US Recommended Uniform Screening Panel (RUSP) [1].

Although there was a slight decrease of participation for some programs in 2020 due to the COVID-19 pandemic, the NSQAP as a whole continued to grow. In 2020, 650 newborn screening laboratories in 88 countries (at least one laboratory per country) participated in the program (Figure 1). Of these laboratories, 484 participated in PT (Table 1) and 383 in QC (Table 2). The program distributed DBS materials for 75 analytes to participating laboratories (Tables 1 and 2).

To offer more specialized services, NSQAP works with the Biochemical Mass Spectrometry Laboratory (BMSL) and the Molecular Quality Improvement Program (MQIP) in the NSMBB to produce and distribute DBS materials.

NSQAP provides quality assurance materials for T4, TSH, 17OHP, IRT, sickle cell and other hemoglobinopathies (Hb), anti-HIV-1 Antibodies (HIV), anti-Toxoplasma Antibodies (TOXO), and the second-tier congenital adrenal hyperplasia (CAH) programs.

BMSL offers newborn screening tandem mass spectrometry (MS/MS) quality assurance, education, and research opportunities. It also oversees the amino acids, acylcarnitines, adrenoleukodystrophy (ALD), BIOT, TGAL, GALT, G6PD, LSD, and filter paper evaluation programs. BMSL also provides second-tier QC programs for maple syrup urine disease/phenylketonuria and homocystinuria.


MQIP oversees the cystic fibrosis DNA PT (CFDNA), T-cell receptor excision circle PT (TREC), and pilot SMA PT programs and provides molecular assay technical assistance to NSQAP participants. In addition, MQIP offers the Molecular Assessment Program (MAP) to US newborn screening laboratories. A MAP visit assesses components of molecular



testing which includes guidance for laboratory-specific needs and assistance in evaluating ongoing and future molecular testing procedures. In-person MAP site visits are currently on hold due to COVID-19 pandemic travel restrictions, however MQIP is offering virtual MAP site visits tailored for specific issues. Contact [CGreene@cdc.gov](mailto:CGreene@cdc.gov) for more information.

**Figure 1.** Eighty-eight countries participated in the Newborn Screening Quality Assurance Program.

 Countries Shown on World Map that Participated in NSQAP During 2020



- |                |             |             |                 |                      |
|----------------|-------------|-------------|-----------------|----------------------|
| Argentina      | Ecuador     | Israel      | New Zealand     | South Korea          |
| Armenia        | Egypt       | Italy       | Nigeria         | Spain                |
| Australia      | El Salvador | Japan       | Norway          | Sri Lanka            |
| Austria        | Estonia     | Jordan      | Oman            | Sweden               |
| Bahrain        | Finland     | Kazakhstan  | Pakistan        | Switzerland          |
| Belgium        | France      | Kuwait      | Panama          | Taiwan               |
| Bolivia        | Georgia     | Laos        | Paraguay        | Tanzania             |
| Brazil         | Germany     | Latvia      | Peru            | Thailand             |
| Bulgaria       | Greece      | Lebanon     | Philippines     | Tunisia              |
| Canada         | Guatemala   | Lithuania   | Poland          | Turkey               |
| Chile          | Honduras    | Luxembourg  | Portugal        | Ukraine              |
| China          | Hong Kong   | Macedonia   | Qatar           | United Arab Emirates |
| Colombia       | Hungary     | Malaysia    | Romania         | United Kingdom       |
| Costa Rica     | Iceland     | Malta       | Saudi Arabia    | United States        |
| Croatia        | India       | Mexico      | Singapore       | Uruguay              |
| Cuba           | Indonesia   | Mongolia    | Slovak Republic | Vietnam              |
| Czech Republic | Iraq        | Morocco     | Slovenia        |                      |
| Denmark        | Ireland     | Netherlands | South Africa    |                      |

**Table 1.** Number of participants reporting proficiency testing analytes. (N = 484)

Note: A "2" after an analyte indicates 2nd tier

Analyte	Total PT Participation in 2020	Analyte	Total PT Participation in 2020
170HP	282	C6	305
T4	78	C8	332
TSH	344	C10	321
TGal	183	C10:1	285
BIOT	213	C10:2	198
GALT	145	C14	303
IRT	237	C14:1	311
G6PD	100	C16	311
CFDNA	73	C160H	308
HGB	79	C18	290
Anti-HIV-1	24	C18:1	284
TOXO	17	C180H	252
TREC	72	170HP2	29
Arg	273	4AD2	29
Cit	300	CORT2	29
Leu	333	11D2	21
Met	315	21D2	22
Phe	419	GALC	11
SUAC	169	GAA	23
Tyr	335	IDUA	24
Val	300	24-LPC	9
C0(L)	318	26-LPC	18
C2(L)	222		
C3	321		
C3DC	117		
C3DC+C40H	140		
C4	296		
C40H	107		
C5	327		
C5:1	290		
C5DC	315		
C5OH	282		

**Table 2.** Number of participants reporting quality control analytes, 2020 (N = 383)

Note: A "2" after an analyte indicates 2nd tier

Analyte	Total QC participation in 2020	Analyte	Total QC participation in 2020
170HP	202	C160H	234
T4	53	C18	233
TSH	280	C180H	196
TGal	129	170HP2	29
GALT	73	4AD2	28
IRT	165	CORT2	28
Ala	196	11D2	22
Arg	209	21D2	22
Cit	231	GALC	20
Gly	171	GAA	40
Leu	246	IDUA	39
Met	237	GLA	29
Orn	177	ABG	28
Phe	324	ASM	17
SUAC	122	20-LPC	23
Tyr	249	22-LPC	23
Val	231	24-LPC	33
C0	239	26-LPC	40
C2	232	GUAC	11
C3	238	CRE2	7
C3DC	82	ALE2	21
C3DC+C40H	123	ILE2	24
C4	232	LEU2	24
C40H	73	PHE2	25
C5	241	TYR2	24
C5:1	205	VAL2	25
C5DC	226	MMA2	31
C5OH	211	EMA2	9
C6	237	MCA2	22
C8	244	tHCY2	30
C10	242	MA2	2
C12	229		
C14	236		
C14:1	214		
C16	238		



## Filter Paper

NSQAP evaluates absorption characteristics of all filter paper lots approved by the Food and Drug Administration (FDA) as a newborn screening collection device [2]. Filter paper manufacturers must establish their own parallel evaluation. NSQAP's evaluations are an impartial and voluntary service offered as a function of our QC program; they do not constitute endorsement of any product.

The disk punched from a DBS specimen gives a volumetric measurement that requires a high degree of uniformity among and within production lots. NSQAP uses an isotopic method developed at CDC to evaluate and compare filter paper lots. It equates mean counts per minute of added radioisotope-labeled T4 contained within a 3.2-mm disk with the serum absorption volume of the disks made from blood with washed, intact red blood cells (RBCs). The latest version of Clinical Laboratory Standards Institute (CLSI) Standard NBS01-A6, Blood Collection on Filter Paper for Newborn Screening Programs, describes the method.

FDA-approved newborn screening filter paper manufacturers (Cytiva Life Sciences and PerkinElmer Health Sciences) provide NSQAP with statistically valid sample sets of unprinted filter paper from each production lot. Tables 3 and 4 show serum absorption volumes from the 10 most recent lots of these two filter paper sources. The published standardized acceptable serum absorption volume per 3.2-mm disk (mean value and 95% confidence interval) is  $1.44 \pm 0.20 \mu\text{L}$ , using blood with washed intact RBCs [2]. The testing results in Tables 3 and 4 are informational only. Each mean value is within the acceptable range for the matrix used. All lots are homogenous (i.e., the measured within-spot, within-sheet, and among-sheets variances were within acceptable limits). CDC used 903™ filter paper lots W171 and W181 to produce the QC and PT specimens distributed in 2020.

**Table 3.** PerkinElmer 226 specimen collection filter paper absorption characteristics by lot number—intact red cells

Filter Paper Lot No.	Date of Evaluation Month/Year	Serum Volume ( $\mu\text{L}$ ) per 3.2 mm (1/8") Punch	Absorption Time (sec)	Spot Diameter (mm)
		Average (StDev)	Average (StDev)	Average (StDev)
114068	Aug 2020	1.44 (0.09)	13.2 (3.8)	16.1 (0.4)
112911	June 2019	1.49 (0.16)	8.4 (1.1)	15.8 (0.7)
112147	Sept 2018	1.49 (0.11)	7.9 (0.9)	15.8 (0.6)
111064	July 2017	1.47 (0.20)	8.2 (1.0)	15.7 (0.5)
110092	July 2016	1.45 (0.09)	9.0 (1.2)	16.0 (0.7)
105617	May 2016	1.46 (0.08)	8.3 (1.8)	15.8 (0.5)
105616	Jan 2016	1.56 (0.11)	10.6 (2.0)	15.6 (0.5)
105178	Aug 2015	1.46 (0.09)	7.8 (1.1)	15.9 (0.6)
104568	March 2015	1.56 (0.10)	10.1 (2.1)	15.9 (0.7)
103649	March 2015	1.53 (0.10)	9.7 (3.1)	15.7 (0.7)

**Table 4.** 903™ specimen collection filter paper absorption characteristics by lot number—intact red cells

Filter Paper Lot No.	Date of Evaluation Month/Year	Serum Volume (µL) per 3.2 mm (1/8") Punch Average (StDev)	Absorption Time (sec) Average (StDev)	Spot Diameter (mm) Average (StDev)
W201	Aug 2020	1.40 (0.09)	14.6 (2.8)	16.1 (0.6)
W191	Oct 2019	1.43 (0.18)	12.2 (2.2)	16.0 (0.7)
W181	Sept 2018	1.42 (0.12)	16.1 (3.3)	16.2 (0.6)
W171	April 2017	1.39 (0.10)	19.7 (4.7)	16.0 (0.7)
W162	Jan 2017	1.43 (0.08)	12.9 (2.7)	16.0 (0.7)
W161	May 2016	1.41 (0.08)	14.8 (3.7)	16.2 (0.8)
W152	Aug 2015	1.37 (0.09)	15.8 (2.4)	16.2 (0.6)
W151	Aug 2015	1.39 (0.08)	15.2 (2.6)	16.2 (0.8)
W142	April 2015	1.46 (0.08)	11.0 (2.2)	16.0 (0.7)
W141	March 2014	1.53 (0.10)	13.8 (3.6)	15.9 (0.6)

## Proficiency Testing

NSQAP usually distributes PT panels three times per year, however due to limitations from the COVID-19 pandemic, two PT events were conducted during 2020. PT panels consisted of five blind-coded specimens. Specimen sets were packaged in a zip-closed, metalized plastic bag with

desiccant. Instructions for analysis and reporting data were located online in the NSQAP portal at <https://nbs.dynamics365portals.us/>. These specimens provided an independent, external assessment of each laboratory's performance.

### The Proficiency Testing Analytes

#### AMINO ACIDS

- arginine (Arg)
- citrulline (Cit)
- leucine (Leu)
- methionine (Met)
- phenylalanine (Phe)
- succinylacetone (SUAC)
- tyrosine (Tyr)
- valine (Val)

#### ACYLCARNITINES

- low free carnitine (C0(L))
- low acetylcarnitine (C2(L))
- propionylcarnitine (C3)
- malonylcarnitine [derivatized] (C3DC)
- C3DC+C40H [non-derivatized]
- butyrylcarnitine (C4)
- hydroxybutyrylcarnitine [derivatized] (C40H)
- tiglylcarnitine (5:1)
- glutaryl carnitine (C5DC)

- hydroxyisovalerylcarnitine (C50H)
- hexanoylcarnitine (C6)
- octanoylcarnitine (C8)
- decanoylcarnitine (C10)
- decenoylcarnitine (C10:1)
- decadienoylcarnitine (C10:2)
- myristoylcarnitine (C14)
- tetradecenoylcarnitine (C14:1)
- palmitoylcarnitine (C16)
- hydroxypalmitoylcarnitine (C160H)
- stearoylcarnitine (C18)
- oleoylcarnitine (C18:1)
- hydroxystearoylcarnitine (C160H)
- stearoylcarnitine (C18)
- oleoylcarnitine (C18:1)
- hydroxystearoylcarnitine (C180H)

#### OTHER ANALYTES

- 17 α-hydroxyprogesterone (17OHP)
- 24:0-lysophosphatidylcholine (C24-LPC)
- 26:0-lysophosphatidylcholine (C26-LPC)
- acid α-glucosidase (GAA)
- α-galactosidase (GLA)
- α-L-iduronidase (IDUA)
- anti-HIV-1 antibodies (HIV)
- anti-toxoplasma antibodies (TOXO)
- biotinidase (BIOT)
- cystic fibrosis DNA variant detection (CFDNA)
- galactoceramidase (GALC)
- galactose-1-phosphate uridylyltransferase (GALT)
- glucose-6-phosphate dehydrogenase (G6PD)
- immunoreactive trypsinogen (IRT)
- second-tier 11-deoxycortisol (11D2)
- second-tier 17 α-hydroxyprogesterone (17OHP2)
- second-tier 21-deoxycortisol (21D2)
- second-tier 4-androstenedione (4AD2)
- second-tier cortisol (CORT2)
- sickle cell disease and other hemoglobinopathies (Hb)
- T-cell receptor excision circle (TREC)
- thyroid-stimulating hormone (TSH)
- thyroxine (T4)
- total galactose (TGAL)

## Proficiency Testing Materials and Methods

NSQAP certified PT specimens for homogeneity, accuracy, stability, and suitability for newborn screening assays. Most PT specimens were prepared from whole blood of 50% hematocrit. PT materials were produced from one of the following: unaltered donor blood, enriched single blood units, or pooled blood units.

**Purified analytes** were used for PT enrichments. Enrichments made with commercially available or custom-synthesized analytes were based on weight. Small variances in enrichments and recoveries might result from impurities in the purchased (synthesized) materials and endogenous analyte concentrations.

**Congenital hypothyroid PT specimens** were enriched with measured amounts of T4 and TSH after reconstituting washed RBCs with purchased T4-depleted charcoal-stripped serum.

**IRT PT specimens** were made from washed, hematocrit-adjusted blood that was treated with a protease inhibitor then enriched with commercially purchased IRT.

**TGal PT specimens** were enriched with galactose and galactose-1-phosphate, allowing measurement of free galactose (galactose alone) and total galactose (free galactose plus galactose-1-phosphate).

**BIOT PT deficient specimens** were made using heat-treated serum combined with compatible donor RBCs.

**GALT and G6PD deficient PT specimens** were made using a 50/50 saline/serum solution combined with compatible washed RBCs, and followed by heat-treatment.

**C0(L) and C2(L) deficient PT specimens** were produced by washing fresh RBCs at least six times then combining with charcoal-stripped serum.

**CFDNA PT specimens** were prepared using blood from anonymous cystic fibrosis patients, carriers, or unaffected individuals without hematocrit adjustment.

**Hb PT specimens** were made from hematocrit-adjusted individual umbilical cord blood units.

**HIV PT DBS specimens** were prepared by mixing purchased donor serum reactive for HIV-1 antibodies and washed RBCs to achieve the desired reactivity.

**TREC PT specimens** were prepared from human blood, including cord blood from unaffected persons and modified adult blood depleted of mononuclear cells or leukocytes.

**LSD PT specimens** were prepared from human blood, including cord blood from unaffected persons and leuko-depleted adult blood restored with lymphoblast cell lines derived from patients with LSD.

**Anti-Toxoplasma Antibody (TOXO) PT DBS specimens** were prepared by combining human serum samples collected from patients exposed to *Toxoplasma gondii* with compatible washed RBCs.

## Proficiency Testing Data Handling

Participants submitted PT data and clinical assessments using the new NSQAP Data reporting portal <https://nbs.dynamics365portals.us/>. Laboratories that submitted results before the data reporting deadline received an individual laboratory evaluation, and their data were included in the data summary report.

## Proficiency Testing Errors

Specimens were evaluated as “Acceptable” or “Unacceptable.” For each analyte and specimen, the participating laboratory’s presumptive clinical assessment must match the CDC certified clinical assessment to achieve an “Acceptable” evaluation. When clinical assessments differed, the evaluation was “Unacceptable”. NSQAP did not identify “Unacceptable” results as “false negative” or “false positive”. It was the responsibility of the laboratory to categorize “Unacceptable” results according to their protocols and policies.

A consensus of 80% of US laboratories, as long as 10 or more US laboratories report results, must be reached for a specimen to be evaluated. If there were less than 10 US laboratories reporting results for any one specimen, all submitted results are evaluated. NSQAP occasionally challenged cutoff levels by enriching samples in the cutoff range and are closely reviewed by the NSQAP PT committee. Not-evaluated specimens were considered educational.

Tables 5–8 show the unacceptable assessments reported in 2020 by domestic and international laboratories by disorder/analyte. The rates for unacceptable misclassifications were based on the total number of specimens tested. Not-evaluated specimens were not included in the error calculations.


The CFDNA PT program provides evaluations based on allele identification and clinical assessment. Allele detection was dependent on the method used. Table 9 summarizes the CFDNA variant challenges distributed in 2020. Table 10 shows the challenges distributed in 2020 for sickle cell disease and other hemoglobinopathies. Participants were evaluated on reported hemoglobin phenotypes and the ability to provide correct clinical assessments.

**Table 5.** Summary of disease specific and non-MSMS proficiency testing errors by domestic laboratories

Analyte/ Disorders	Specimens assayed (N)	Unacceptable Assessments (%)
<b>anti-HIV-1 Antibodies</b>	65	0.0%
<b>anti-Toxoplasma Antibodies</b>	20	0.0%
<b>Congenital Adrenal Hyperplasia</b>	405	1.7%
<b>Biotinidase Deficiency</b>	405	2.5%
<b>G6PD Deficiency</b>	25	0.0%
<b>GALT Deficiency</b>	420	1.0%
<b>Immunoreactive Trypsinogen</b>	420	1.2%
<b>Congenital Hypothyroidism</b>	415	0.0%
<b>Galactosemia</b>	200	1.0%
<b>Lysosomal Storage Disorder Krabbe</b>	105	0.0%
<b>Lysosomal Storage Disorder Pompe</b>	215	0.5%
<b>Lysosomal Storage Disorder Mucopolysaccharidosis Type 1</b>	220	0.0%
<b>T-Cell Receptor Excision Circle</b>	287	0.7%
<b>Second-tier Congenital Adrenal Hyperplasia</b>	77	6.5%
<b>ALD 24:0 Lysophosphatidylcholine</b>	65	1.5%
<b>ALD 26:0 Lysophosphatidylcholine</b>	160	0.0%
<b>Sickle Cell and Other Hemoglobinopathies Phenotype Errors</b>	435	0.5%
<b>Sickle Cell and Other Hemoglobinopathies Clinical Assessment Errors</b>	435	0.5%
<b>Cystic Fibrosis DNA Variant Allele Errors</b>	337	0.3%
<b>Cystic Fibrosis DNA Variant Clinical Assessment Errors</b>	337	0.0%

**Table 6.** Summary of disease specific and non-MSMS proficiency testing errors by international laboratories

Analyte/ Disorders	Specimens Assayed (N)	Unacceptable Assessments (%)
<b>anti-HIV-1 Antibodies</b>	100	1.0%
<b>anti-Toxoplasma Antibodies</b>	55	7.3%
<b>Congenital Adrenal Hyperplasia</b>	1780	0.7%
<b>Biotinidase Deficiency</b>	1315	2.4%
<b>G6PD Deficiency</b>	670	1.6%
<b>GALT Deficiency</b>	850	1.3%
<b>Immunoreactive Trypsinogen</b>	1495	1.9%
<b>Congenital Hypothyroidism</b>	2190	0.8%
<b>Galactosemia</b>	1215	1.6%
<b>T-Cell Receptor Excision Circle</b>	194	2.1%
<b>Second-tier Congenital Adrenal Hyperplasia</b>	186	7.5%
<b>ALD 24:0 Lysophosphatidylcholine</b>	70	4.3
<b>ALD 26:0 Lysophosphatidylcholine</b>	95	5.3
<b>Sickle Cell and Other Hemoglobinopathies Phenotype Errors</b>	345	3.2%
<b>Sickle Cell and Other Hemoglobinopathies Clinical Assessment Errors</b>	345	2.4%
<b>Cystic Fibrosis DNA Variant Allele Errors</b>	190	0.0%
<b>Cystic Fibrosis DNA Variant Clinical Assessment Errors</b>	190	0.0%



**Table 7.** Summary of amino acid and acylcarnitine proficiency test errors by domestic laboratories

Analyte	Specimens Assayed (N)	Unacceptable Assessments (%)
<b>Arginine Screen</b>	345	0.6%
<b>Citrulline Screen</b>	435	1.8%
<b>Leucine Screen</b>	435	2.8%
<b>Methionine Screen</b>	425	0.5%
<b>Phenylalanine Screen</b>	510	0.8%
<b>Succinylacetone Screen</b>	365	0.0%
<b>Tyrosine Screen</b>	455	1.8%
<b>Valine Screen</b>	305	1.6%
<b>C0(L) Screen</b>	455	0.7%
<b>C2(L) Screen</b>	225	5.3%
<b>C3 Screen</b>	455	0.4%
<b>C3DC Screen</b>	120	0.8%
<b>C3DC+C40H Screen</b>	240	0.0%
<b>C4 Screen</b>	400	0.8%
<b>C40H Screen</b>	110	0.0%
<b>C5 Screen</b>	455	0.0%
<b>C5:1 Screen</b>	445	2.0%
<b>C5DC Screen</b>	445	0.4%
<b>C50H Screen</b>	445	1.1%
<b>C6 Screen</b>	415	0.7%
<b>C8 Screen</b>	455	0.2%
<b>C10 Screen</b>	415	0.0%
<b>C10:1 Screen</b>	385	0.3%
<b>C10:2 Screen</b>	265	3.0%
<b>C14 Screen</b>	405	0.7%
<b>C14:1 Screen</b>	455	0.4%
<b>C16 Screen</b>	425	0.0%
<b>C160H Screen</b>	455	0.0%
<b>C18 Screen</b>	370	3.5%
<b>C18:1 Screen</b>	395	1.8%
<b>C180H Screen</b>	350	0.9%

**Table 8.** Summary of amino acid and acylcarnitine proficiency testing errors by international laboratories


Analyte	Specimens Assayed (N)	Unacceptable Assessments (%)
<b>Arginine Screen</b>	1740	1.2%
<b>Citrulline Screen</b>	990	0.4%
<b>Leucine Screen</b>	2180	1.4%
<b>Methionine Screen</b>	2040	2.5%
<b>Phenylalanine Screen</b>	2680	1.7%
<b>Succinylacetone Screen</b>	985	0.4%
<b>Tyrosine Screen</b>	2135	0.6%
<b>Valine Screen</b>	2010	1.5%
<b>C0(L) Screen</b>	2065	2.6%
<b>C2(L) Screen</b>	1435	3.2%
<b>C3 Screen</b>	2070	1.6%
<b>C3DC Screen</b>	770	1.9%
<b>C3DC+C40H Screen</b>	740	0.7%
<b>C4 Screen</b>	1900	0.3%
<b>C40H Screen</b>	680	2.6%
<b>C5 Screen</b>	2125	0.9%
<b>C5:1 Screen</b>	1830	1.3%
<b>C5DC Screen</b>	2050	1.3%
<b>C50H Screen</b>	1810	6.8%
<b>C6 Screen</b>	1955	1.4%
<b>C8 Screen</b>	2155	0.9%
<b>C10 Screen</b>	2090	0.8%
<b>C10:1 Screen</b>	1830	1.6%
<b>C10:2 Screen</b>	1220	2.5%
<b>C14 Screen</b>	1940	0.8%
<b>C14:1 Screen</b>	2005	0.9%
<b>C16 Screen</b>	1995	0.4%
<b>C160H Screen</b>	1990	0.8%
<b>C18 Screen</b>	1875	2.5%
<b>C18:1 Screen</b>	1825	0.9%
<b>C180H Screen</b>	1590	0.8%

**Table 9.** Cystic Fibrosis DNA variant (CFTR gene) challenges distributed in 2020

Variant (Legacy Name)	Variant (HGVS Nomenclature)	Variants Sent
<b>F508del</b>	(c.1521_1523delCTT)	5
<b>1717-1G&gt;A</b>	(c.1585-1G>A)	1
<b>2105-2117del13insAGAAA</b>	(c.1973_1985del13insAGAAA)	1
<b>2183AA&gt;G</b>	(c.2051_2052delAAinsG)	1
<b>3849+10kb C&gt;T</b>	(c.3717+12191C>T)	1
<b>D1152H</b>	(c.3454G>C)	1
<b>N1303K</b>	(c.3909C>G)	1
<b>R347P</b>	(c.1040G>C)	1
<b>Y1092X</b>	(c.3276C>A or c.3276C>G)	1
<b>Wild type</b>	Wild type	7

**Table 10.** Hemoglobinopathies accepted presumptive phenotype distributed in 2020

Quarter	Specimen 1	Specimen 2	Specimen 3	Specimen 4	Specimen 5
<b>Quarter 1</b>	FAC	FA	FS	FAS	FAS
<b>Quarter 4</b>	FAS	FAC	FA	FAS	FAC



## Proficiency Testing Cutoff Values

Participants reported the decision level for sorting test results as presumptive positive (outside normal limits) from results reported as negative (within normal limits), based on their established cutoff value. Because CDC does not test newborns, establishing a population cutoff value is not possible. Therefore, CDC cutoff values are determined by using the mean of all domestic laboratory cutoff values. (Note: Each laboratory should establish its own cutoff values rather than using the CDC-reported cutoff values.)

Tables 11–14 summarize the reported cutoff values for domestic and international laboratories. The tables show summary statistics for each analyte. Tables 15–17 summarize domestic cutoff statistics by method.

**Table 11.** Summary of non-MS/MS cutoff values for domestic laboratories

Analyte	N	Mean	Median	Mode	Minimum	Maximum
<b>170HP (ng/mL serum)</b>	40	34.7	33.0	30.0	20.0	70.0
<b>IRT (ng/mL blood)</b>	40	61.2	60.0	65.0	39.7	100.0
<b>T4 (µg/dL serum)</b>	18	6.0	6.0	5.0	5.0	8.0
<b>TGal (mg/dL blood)</b>	17	10.7	10.0	10.0	6.0	15.0
<b>TSH (µIU/mL serum)</b>	41	30.5	25.5	20.0	10.0	58.0

**Table 12.** Summary of non-MS/MS cutoff values for international laboratories

Analyte	N	Mean	Median	Mode	Minimum	Maximum
<b>170HP (ng/mL serum)</b>	175	21.4	18.3	35.0	5.2	60.0
<b>IRT (ng/mL blood)</b>	137	64.2	63.0	60.0	20.0	90.8
<b>T4 (µg/dL serum)</b>	31	7.2	6.0	6.0	1.9	22.0
<b>TGal (mg/dL blood)</b>	110	16.5	10.0	10.0	1.5	30.0
<b>TSH (µIU/mL serum)</b>	213	21.9	20.0	20.0	4.5	44.4
<b>Phe (µmol/L blood)</b>	58	157.7	136.7	121.2	103.0	303.0



**Table 13.** Summary of amino acid and acylcarnitine cutoff values for domestic laboratories ( $\mu\text{mol/L}$  blood)

Analyte	N	Mean	Median	Mode	Minimum	Maximum
<b>Arginine Screen</b>	33	74.3	70.0	50.0	35.0	120.0
<b>Citrulline Screen</b>	43	58.2	55.0	60.0	34.0	200.0
<b>Leucine Screen</b>	43	287.9	275.0	250.0	145.0	400.0
<b>Methionine Screen</b>	42	73.6	75.0	100.0	40.0	100.0
<b>Phenylalanine Screen</b>	50	141.7	150.0	130.0	75.0	188.0
<b>Succinylacetone Screen</b>	37	2.3	1.9	1.0	0.8	5.4
<b>Tyrosine Screen</b>	45	414.1	360.0	300.0	92.0	850.0
<b>Valine Screen</b>	30	291.5	291.0	250.0	180.0	530.0
<b>C0(L) Screen</b>	44	8.21	7.00	6.00	5.00	24.00
<b>C2(L) Screen</b>	23	6.54	7.00	9.00	0.00	10.00
<b>C3 Screen</b>	45	5.72	6.00	6.30	3.10	8.00
<b>C3DC Screen</b>	10	0.24	0.22	0.22	0.10	0.45
<b>C3DC+C40H Screen</b>	24	0.55	0.41	0.38	0.25	3.03
<b>C4 Screen</b>	39	1.26	1.27	1.70	0.49	1.90
<b>C40H Screen</b>	9	0.63	0.70	0.70	0.20	0.80
<b>C5 Screen</b>	45	0.70	0.62	0.60	0.34	1.20
<b>C5:1 Screen</b>	44	0.19	0.13	0.10	0.02	0.50
<b>C5DC Screen</b>	44	0.38	0.43	0.50	0.05	0.80
<b>C50H Screen</b>	44	0.85	0.84	0.80	0.32	1.50
<b>C6 Screen</b>	41	0.38	0.26	0.24	0.14	0.95
<b>C8 Screen</b>	45	0.44	0.43	0.60	0.12	0.73
<b>C10 Screen</b>	41	0.44	0.40	0.40	0.20	0.70
<b>C10:1 Screen</b>	38	0.27	0.25	0.25	0.12	0.45
<b>C10:2 Screen</b>	26	0.14	0.10	0.10	0.02	0.38
<b>C14 Screen</b>	40	0.72	0.70	0.70	0.27	1.20
<b>C14:1 Screen</b>	45	0.60	0.60	0.60	0.17	0.80
<b>C16 Screen</b>	42	7.80	7.85	10.00	2.14	10.36
<b>C160H Screen</b>	45	0.12	0.10	0.10	0.06	0.25
<b>C18 Screen</b>	35	2.33	2.20	3.50	0.70	3.50
<b>C18:1 Screen</b>	39	3.59	3.00	3.00	2.00	7.00
<b>C180H Screen</b>	34	0.09	0.10	0.10	0.03	0.18

**Table 14.** Summary of amino acid and acylcarnitine cutoff values for international laboratories ( $\mu\text{mol/L}$  blood)

Analyte	N	Mean	Median	Mode	Minimum	Maximum
<b>Arginine Screen</b>	166	57.6	57.7	70.0	10.0	150.0
<b>Citrulline Screen</b>	181	49.8	49.1	55.0	20.0	100.0
<b>Leucine Screen</b>	205	308.4	294.0	300.0	112.0	686.7
<b>Methionine Screen</b>	193	55.0	50.0	75.0	20.0	140.0
<b>Phenylalanine Screen</b>	199	130.2	120.0	150.0	48.0	300.0
<b>Succinylacetone Screen</b>	90	2.2	1.7	2.0	0.3	8.0
<b>Tyrosine Screen</b>	200	296.3	290.0	350.0	79.9	600.0
<b>Valine Screen</b>	191	270.2	271.4	300.0	142.0	470.0
<b>C0(L) Screen</b>	185	8.37	8.00	8.00	1.70	24.70
<b>C2(L) Screen</b>	110	7.03	7.00	7.00	0.00	21.20
<b>C3 Screen</b>	197	5.31	5.20	5.65	1.30	11.02
<b>C3DC Screen</b>	69	0.27	0.25	0.25	0.04	1.40
<b>C3DC+C40H Screen</b>	72	0.49	0.45	0.45	0.15	2.53
<b>C4 Screen</b>	183	0.95	0.97	1.30	0.16	2.50
<b>C40H Screen</b>	61	0.60	0.61	0.65	0.05	1.20
<b>C5 Screen</b>	201	0.67	0.60	0.70	0.10	2.00
<b>C5:1 Screen</b>	176	0.15	0.11	0.25	0.01	0.90
<b>C5DC Screen</b>	195	0.34	0.30	0.35	0.03	1.00
<b>C50H Screen</b>	172	0.71	0.74	0.80	0.15	1.50
<b>C6 Screen</b>	184	0.27	0.20	0.40	0.04	1.30
<b>C8 Screen</b>	206	0.34	0.30	0.45	0.05	1.30
<b>C10 Screen</b>	193	0.36	0.35	0.45	0.07	1.10
<b>C10:1 Screen</b>	172	0.25	0.21	0.30	0.06	1.00
<b>C10:2 Screen</b>	123	0.15	0.10	0.15	0.01	2.00
<b>C14 Screen</b>	182	0.60	0.56	0.75	0.08	1.30
<b>C14:1 Screen</b>	191	0.46	0.40	0.60	0.04	2.50
<b>C16 Screen</b>	187	6.92	7.00	7.50	0.57	14.00
<b>C160H Screen</b>	189	0.11	0.10	0.10	0.02	0.75
<b>C18 Screen</b>	179	2.10	2.00	2.30	0.10	9.00
<b>C18:1 Screen</b>	176	3.09	3.00	3.50	0.15	7.00
<b>C180H Screen</b>	153	0.08	0.07	0.10	0.01	0.50

**Table 15.** Summary of cutoff values by analyte and method for domestic laboratories — hormones, galactose, and immunoreactive trypsinogen, (methods N<3 not shown)

### 17 $\alpha$ -Hydroxyprogesterone ng/mL serum

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>40</b>	<b>34.7</b>	<b>33.0</b>	<b>30.0</b>	<b>20.0</b>	<b>70.0</b>
AutoDELFIA® Neonatal 17OHP PerkinElmer	13	37.3	33.0	33.0	25.0	70.0
GSP® TGal Neonatal PerkinElmer	26	33.5	31.0	30.0	20.0	55.0

### Immunoreactive Trypsinogen ng/mL blood

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>40</b>	<b>61.2</b>	<b>60.0</b>	<b>65.0</b>	<b>39.7</b>	<b>100.0</b>
AutoDELFIA® Neonatal IRT PerkinElmer	19	67.4	65.0	65.0	51.0	100.0
GSP® TGal Neonatal PerkinElmer	21	55.5	55.0	55.0	39.7	70.0

### Thyroxine $\mu$ g/dL serum

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>18</b>	<b>6.0</b>	<b>6.0</b>	<b>5.0</b>	<b>5.0</b>	<b>8.0</b>
AutoDELFIA® Neonatal T4 PerkinElmer	4	6.0	6.3	n/a	5.0	6.6
GSP® TGal Neonatal PerkinElmer	13	6.0	5.5	5.0	5.0	8.0

### Thyroid-Stimulating Hormone $\mu$ U/mL serum

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>41</b>	<b>30.5</b>	<b>25.5</b>	<b>20.0</b>	<b>10.0</b>	<b>58.0</b>
AutoDELFIA® Neonatal hTSH PerkinElmer	15	39.4	36.0	20.0	20.0	58.0
GSP® TGal Neonatal PerkinElmer	25	25.6	25.0	20.0	10.0	54.0

### Total Galactose mg/dL blood

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>17</b>	<b>10.7</b>	<b>10.0</b>	<b>10.0</b>	<b>6.0</b>	<b>15.0</b>
50hr Reagent Kit Spotcheck® TGal Astoria-Pacific	3	12.0	11.0	n/a	10.0	15.0
GSP® TGal Neonatal PerkinElmer	10	11.2	11.0	10.0	7.3	14.0



**Table 16.** Domestic cutoff summary by analyte and method—amino acids (µmol/L blood), (methods N < 3 not shown)

### Arginine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>33</b>	<b>74.3</b>	<b>70.0</b>	<b>50.0</b>	<b>35.0</b>	<b>120.0</b>
Derivatized - MS/MS non-kit	6	64.7	65.0	n/a	35.0	100.0
Non-derivatized - MS/MS NeoBase™ PerkinElmer	17	76.5	80.0	50.0	48.0	120.0
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	6	80.8	92.5	50.0	50.0	105.0

### Citrulline

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>43</b>	<b>58.2</b>	<b>55.0</b>	<b>60.0</b>	<b>34.0</b>	<b>200.0</b>
Derivatized - MS/MS non-kit	9	57.2	55.0	55.0	34.0	75.0
Non-derivatized - MS/MS NeoBase™ PerkinElmer	21	58.0	60.0	60.0	40.0	75.0
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	49.0	50.0	50.0	45.0	56.0
Non-derivatized - MS/MS non-kit	4	84.8	52.0	n/a	35.0	200.0

### Leucine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>43</b>	<b>287.9</b>	<b>275.0</b>	<b>250.0</b>	<b>145.0</b>	<b>400.0</b>
Derivatized - MS/MS non-kit	9	280.7	289.0	220.0	220.0	400.0
Non-derivatized - MS/MS NeoBase™ PerkinElmer	21	298.4	275.0	250.0	225.0	400.0
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	275.7	250.0	n/a	145.0	378.0
Non-derivatized - MS/MS non-kit	3	285.0	300.0	n/a	250.0	305.0

### Methionine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>42</b>	<b>73.6</b>	<b>75.0</b>	<b>100.0</b>	<b>40.0</b>	<b>100.0</b>
Derivatized - MS/MS non-kit	9	65.8	67.0	70.0	44.0	83.0
Non-derivatized - MS/MS NeoBase™ PerkinElmer	21	82.0	85.0	100.0	50.0	100.0
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	67.7	74.0	45.0	45.0	90.0
Non-derivatized - MS/MS non-kit	3	51.7	55.0	n/a	40.0	60.0

### Phenylalanine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>50</b>	<b>141.7</b>	<b>150.0</b>	<b>130.0</b>	<b>75.0</b>	<b>188.0</b>
Derivatized - MS/MS non-kit	11	137.5	150.0	150.0	75.0	182.0
Non-derivatized - MS/MS NeoBase™ PerkinElmer	21	148.1	160.0	165.0	120.0	180.0
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	138.1	150.0	n/a	100.0	160.0
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	4	127.5	130.0	130.0	100.0	150.0
Non-derivatized - MS/MS non-kit	6	124.2	125.5	150.0	74.0	150.0

## Succinylacetone

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>37</b>	<b>2.3</b>	<b>1.9</b>	<b>1.0</b>	<b>0.8</b>	<b>5.4</b>
Derivatized - MS/MS non-kit	8	3.1	2.4	2.0	1.9	5.4
Non-derivatized - MS/MS NeoBase™ PerkinElmer	20	2.4	1.8	4.5	0.8	4.5
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	1.5	1.0	1.0	0.9	2.8

## Tyrosine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>45</b>	<b>414.1</b>	<b>360.0</b>	<b>300.0</b>	<b>92.0</b>	<b>850.0</b>
Derivatized - MS/MS non-kit	10	317.2	327.5	300.0	92.0	500.0
Non-derivatized - MS/MS NeoBase™ PerkinElmer	19	536.9	450.0	850.0	300.0	850.0
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	359.6	350.0	350.0	243.0	500.0
Non-derivatized - MS/MS non-kit	4	297.5	290.0	290.0	250.0	360.0

## Valine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>30</b>	<b>291.5</b>	<b>291.0</b>	<b>250.0</b>	<b>180.0</b>	<b>530.0</b>
Derivatized - MS/MS non-kit	7	275.0	240.0	240.0	180.0	420.0
Non-derivatized - MS/MS NeoBase™ PerkinElmer	11	332.0	325.0	300.0	250.0	530.0
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	6	250.3	241.0	180.0	180.0	360.0
Non-derivatized - MS/MS non-kit	3	238.7	250.0	n/a	210.0	256.0

**Table 17.** Domestic cutoff summary by analyte and method—acylcarnitines (μmol/L blood), (methods N < 3 not shown)

## CO(L)

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>44</b>	<b>8.21</b>	<b>7.00</b>	<b>6.00</b>	<b>5.00</b>	<b>24.00</b>
Derivatized - MS/MS non-kit	11	10.82	9.20	10.00	5.60	24.00
Non-derivatized - MS/MS NeoBase™ PerkinElmer	20	6.64	6.00	6.00	5.00	10.00
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	8	6.74	6.65	6.00	6.00	7.50
Non-derivatized - MS/MS non-kit	4	9.38	7.50	n/a	6.50	16.00

## C2(L)

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>23</b>	<b>6.54</b>	<b>7.00</b>	<b>9.00</b>	<b>0.00</b>	<b>10.00</b>
Derivatized - MS/MS non-kit	6	6.72	7.16	n/a	2.00	10.00
Non-derivatized - MS/MS NeoBase™ PerkinElmer	8	6.90	7.00	7.00	4.00	9.00
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	3	7.17	9.00	9.00	3.50	9.00
Non-derivatized - MS/MS non-kit	4	4.83	5.85	n/a	0.00	7.60

### C3

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>45</b>	<b>5.72</b>	<b>6.00</b>	<b>6.30</b>	<b>3.10</b>	<b>8.00</b>
Derivatized - MS/MS non-kit	11	5.10	4.81	n/a	3.30	7.70
Non-derivatized - MS/MS NeoBase™ PerkinElmer	20	5.97	6.05	6.30	4.80	8.00
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	8	6.14	6.25	n/a	4.00	7.55
Non-derivatized - MS/MS non-kit	4	5.63	6.25	n/a	3.10	6.92

### C3DC

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>10</b>	<b>0.24</b>	<b>0.22</b>	<b>0.22</b>	<b>0.10</b>	<b>0.45</b>
Derivatized - MS/MS non-kit	8	0.24	0.22	0.22	0.10	0.45

### C3DC + C4OH

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>24</b>	<b>0.55</b>	<b>0.41</b>	<b>0.38</b>	<b>0.25</b>	<b>3.03</b>
Non-derivatized - MS/MS NeoBase™ PerkinElmer	16	0.39	0.38	0.38	0.25	0.60
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	5	0.47	0.50	0.50	0.41	0.52
Non-derivatized - MS/MS non-kit	3	1.52	1.20	n/a	0.33	3.03

### C4

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>39</b>	<b>1.26</b>	<b>1.27</b>	<b>1.70</b>	<b>0.49</b>	<b>1.90</b>
Derivatized - MS/MS non-kit	8	1.24	1.33	0.80	0.79	1.90
Non-derivatized - MS/MS NeoBase™ PerkinElmer	17	1.35	1.30	1.70	1.00	1.70
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	8	1.30	1.31	1.30	0.90	1.70
Non-derivatized - MS/MS non-kit	4	1.00	1.15	1.20	0.49	1.20

### C4OH

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>9</b>	<b>0.63</b>	<b>0.70</b>	<b>0.70</b>	<b>0.20</b>	<b>0.80</b>
Derivatized - MS/MS non-kit	7	0.63	0.70	0.80	0.20	0.80

### C5

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>45</b>	<b>0.70</b>	<b>0.62</b>	<b>0.60</b>	<b>0.34</b>	<b>1.20</b>
Derivatized - MS/MS non-kit	11	0.74	0.66	0.70	0.34	1.20
Non-derivatized - MS/MS NeoBase™ PerkinElmer	20	0.73	0.68	1.00	0.45	1.00
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	8	0.68	0.60	0.58	0.54	1.00
Non-derivatized - MS/MS non-kit	4	0.55	0.55	n/a	0.40	0.68

## C5:1

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>44</b>	<b>0.19</b>	<b>0.13</b>	<b>0.10</b>	<b>0.02</b>	<b>0.50</b>
Derivatized - MS/MS non-kit	11	0.22	0.15	0.08	0.07	0.50
Non-derivatized - MS/MS NeoBase™ PerkinElmer	20	0.22	0.15	0.10	0.03	0.50
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	8	0.15	0.10	0.10	0.02	0.50
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	4	0.09	0.07	n/a	0.04	0.19
Non-derivatized - MS/MS non-kit	4	0.23	0.19	0.10	0.03	0.50

## C5DC

Method	N	Mean	Median	Mode	Min	Max
<b>All Methods</b>	<b>44</b>	<b>0.38</b>	<b>0.43</b>	<b>0.50</b>	<b>0.05</b>	<b>0.80</b>
Derivatized - MS/MS non-kit	11	0.18	0.15	0.13	0.10	0.30
Non-derivatized - MS/MS NeoBase™ PerkinElmer	20	0.50	0.50	0.50	0.30	0.80
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	0.38	0.45	0.24	0.24	0.51
Non-derivatized - MS/MS non-kit	4	0.35	0.38	n/a	0.05	0.60

## C5OH

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>44</b>	<b>0.85</b>	<b>0.84</b>	<b>0.80</b>	<b>0.32</b>	<b>1.50</b>
Derivatized - MS/MS non-kit	11	0.82	0.80	0.80	0.32	1.36
Non-derivatized - MS/MS NeoBase™ PerkinElmer	20	0.80	0.80	0.85	0.60	1.05
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	1.06	0.92	1.50	0.70	1.50
Non-derivatized - MS/MS non-kit	4	0.93	0.99	n/a	0.52	1.20

## C6

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>41</b>	<b>0.38</b>	<b>0.26</b>	<b>0.24</b>	<b>0.14</b>	<b>0.95</b>
Derivatized - MS/MS non-kit	10	0.35	0.31	n/a	0.14	0.63
Non-derivatized - MS/MS NeoBase™ PerkinElmer	18	0.45	0.25	0.95	0.16	0.95
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	0.36	0.26	0.24	0.16	0.95
Non-derivatized - MS/MS non-kit	4	0.21	0.20	0.15	0.15	0.30

## C8

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>45</b>	<b>0.44</b>	<b>0.43</b>	<b>0.60</b>	<b>0.12</b>	<b>0.73</b>
Derivatized - MS/MS non-kit	11	0.42	0.49	0.50	0.12	0.73
Non-derivatized - MS/MS NeoBase™ PerkinElmer	20	0.48	0.44	0.60	0.32	0.70
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	8	0.46	0.46	0.60	0.29	0.60
Non-derivatized - MS/MS non-kit	4	0.33	0.30	0.30	0.23	0.50

## C10

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>41</b>	<b>0.44</b>	<b>0.40</b>	<b>0.40</b>	<b>0.20</b>	<b>0.70</b>
Derivatized - MS/MS non-kit	10	0.40	0.43	0.50	0.24	0.55
Non-derivatized - MS/MS NeoBase™ PerkinElmer	18	0.49	0.47	0.65	0.30	0.70
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	0.46	0.45	0.55	0.25	0.65
Non-derivatized - MS/MS non-kit	4	0.29	0.28	n/a	0.20	0.40

## C10:1

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>38</b>	<b>0.27</b>	<b>0.25</b>	<b>0.25</b>	<b>0.12</b>	<b>0.45</b>
Derivatized - MS/MS non-kit	9	0.28	0.25	0.25	0.17	0.37
Non-derivatized - MS/MS NeoBase™ PerkinElmer	16	0.30	0.25	0.45	0.15	0.45
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	0.23	0.16	0.13	0.12	0.45
Non-derivatized - MS/MS non-kit	4	0.23	0.19	n/a	0.15	0.40

## C10:2

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>26</b>	<b>0.14</b>	<b>0.10</b>	<b>0.10</b>	<b>0.02</b>	<b>0.38</b>
Derivatized - MS/MS non-kit	7	0.21	0.20	0.10	0.10	0.38
Non-derivatized - MS/MS NeoBase™ PerkinElmer	9	0.10	0.10	0.10	0.04	0.15
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	6	0.12	0.10	0.10	0.02	0.21
Non-derivatized - MS/MS non-kit	3	0.07	0.06	n/a	0.04	0.12

## C14

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>40</b>	<b>0.72</b>	<b>0.70</b>	<b>0.70</b>	<b>0.27</b>	<b>1.20</b>
Derivatized - MS/MS non-kit	10	0.66	0.70	0.70	0.32	0.96
Non-derivatized - MS/MS NeoBase™ PerkinElmer	17	0.82	0.70	0.70	0.46	1.20
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	0.72	0.60	0.60	0.57	1.20
Non-derivatized - MS/MS non-kit	4	0.54	0.55	n/a	0.27	0.80

## C14:1

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>45</b>	<b>0.60</b>	<b>0.60</b>	<b>0.60</b>	<b>0.17</b>	<b>0.80</b>
Derivatized - MS/MS non-kit	11	0.54	0.60	0.60	0.24	0.75
Non-derivatized - MS/MS NeoBase™ PerkinElmer	20	0.68	0.68	0.80	0.50	0.80
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	8	0.61	0.60	0.60	0.50	0.80
Non-derivatized - MS/MS non-kit	4	0.45	0.51	n/a	0.17	0.60

## C16

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>42</b>	<b>7.80</b>	<b>7.85</b>	<b>10.00</b>	<b>2.14</b>	<b>10.36</b>
Derivatized - MS/MS non-kit	10	7.41	7.40	n/a	5.19	10.00
Non-derivatized - MS/MS NeoBase™ PerkinElmer	18	8.29	8.00	10.00	5.00	10.00
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	8	8.33	8.91	6.00	6.00	10.36
Non-derivatized - MS/MS non-kit	4	6.04	6.65	n/a	2.14	8.70

## C160H

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>45</b>	<b>0.12</b>	<b>0.10</b>	<b>0.10</b>	<b>0.06</b>	<b>0.25</b>
Derivatized - MS/MS non-kit	11	0.13	0.11	0.10	0.10	0.25
Non-derivatized - MS/MS NeoBase™ PerkinElmer	20	0.10	0.10	0.10	0.07	0.16
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	8	0.11	0.10	0.10	0.06	0.16
Non-derivatized - MS/MS non-kit	4	0.17	0.16	n/a	0.11	0.25

## C18

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>35</b>	<b>2.33</b>	<b>2.20</b>	<b>3.50</b>	<b>0.70</b>	<b>3.50</b>
Derivatized - MS/MS non-kit	6	1.79	1.75	n/a	1.31	2.50
Non-derivatized - MS/MS NeoBase™ PerkinElmer	17	2.58	2.46	3.50	1.55	3.50
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	6	2.66	2.43	2.30	2.20	3.50
Non-derivatized - MS/MS non-kit	4	1.79	2.00	2.00	0.70	2.47

## C18:1

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>39</b>	<b>3.59</b>	<b>3.00</b>	<b>3.00</b>	<b>2.00</b>	<b>7.00</b>
Derivatized - MS/MS non-kit	9	2.80	3.00	3.00	2.00	3.50
Non-derivatized - MS/MS NeoBase™ PerkinElmer	17	4.34	3.50	7.00	2.00	7.00
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	7	3.41	2.70	2.70	2.50	7.00
Non-derivatized - MS/MS non-kit	4	3.03	3.07	n/a	2.00	4.00

## C180H

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>34</b>	<b>0.09</b>	<b>0.10</b>	<b>0.10</b>	<b>0.03</b>	<b>0.18</b>
Derivatized - MS/MS non-kit	6	0.09	0.09	0.10	0.03	0.18
Non-derivatized - MS/MS NeoBase™ PerkinElmer	16	0.09	0.10	0.10	0.03	0.16
Non-derivatized - MS/MS NeoBase™2 PerkinElmer	6	0.08	0.09	0.04	0.04	0.13
Non-derivatized - MS/MS non-kit	4	0.09	0.11	n/a	0.04	0.12

## 2020 Bias Plots

### Proficiency Testing Bias Plots

Figures 2–37 are illustrated for PT analytes reported using the NSQAP Participant Portal. Bias plots for each analyte were selected to compare PT results by different methods. The CDC expected value (EV) of each specimen equals the sum of the enriched value and the endogenous (non-enriched) value. For IRT PT specimens, the CDC-assayed value is reported.

Non-derivatized MS/MS methods for amino acids and acylcarnitine analysis cannot distinguish between analytes C3DC and C4OH (i.e., they are isobaric). Laboratories using a non-derivatized MS/MS method report C3DC+C4OH, while derivatized MS/MS method users report these analytes separately. The bias plots show the difference of the reported value (positive or negative) by laboratory and method subtracted from the expected or assayed value. To illustrate method-related differences in analyte recoveries, the PT quantitative results are grouped by kit or method.

For each plot, note the scale-changes of the y-axis. A reported value matching the EV falls on the plot's "0" line. For each figure, a summary of the specimen data for the selected PT challenge is tabulated in the left margin. Ideally, a reasonable bias is less than 20% of the EV.

The bias plots illustrate the 95% confidence interval for the participant mean. A tight scatter within this interval indicates good performance for a method or a group of methods. In general, the quantitative comparisons for PT challenges are reasonable within a method but may vary between methods. Because some of the pools in a routine PT survey represent a unique donor specimen, differences in endogenous materials in the donor specimens might influence method-related differences.

Note for accessibility:

For Figures 2–37, the bias plot's explanation follows each figure title.

**Figure 2. Reproducibility of Results:  
Bias Plot of 17  $\alpha$ -Hydroxyprogesterone (17OHP) Values by Method  
Quarter 1, Specimen 20201001004  
Expected Value (EV) = 86.0 mg/mL serum**

**Quarter 1**


Specimen: 20201001004

**Enriched: 85.0**

**CDC Assayed: 78.2**

**Participant Mean: 79.6**

**Participant Bias: -6.4**



The 17OHP bias plot shows units of measure on the y-axis ranging from 100 mg/mL serum to -100 mg/mL serum. The bias for this plot is -6.4 ng/mL serum. The data shows an even scatter among most participants with some results outside the 95% upper and lower bias confidence intervals.

**Figure 3. Reproducibility of Results:  
Bias Plot of Thyroxine (T4) Values by Method  
Quarter 1, Specimen 20201001003  
Expected Value (EV) = 13.1  $\mu$ g/dL serum**

**Quarter 1**


Specimen: 20201001003

**Enriched: 10.0**

**CDC Assayed: 14.2**

**Participant Mean: 12.9**

**Participant Bias : -0.2**



The T4 bias plot shows units of measure on the y-axis ranging from 30  $\mu$ g/dL serum to -30  $\mu$ g/dL serum. The bias for this plot is -0.2. The data shows a good agreement among participants with only a few points above the 95% upper limit.

**Figure 4. Reproducibility of Results:  
Bias Plot of Total Galactose (TGal) Values by Method  
Quarter 1, Specimen 20201001002  
Expected Value (EV) = 26.0 mg/dL blood**

**Quarter 1**


Specimen: 20201001003

**Enriched: 25.0**

**CDC Assayed: 21.7**

**Participant Mean: 24.9**

**Participant Bias: -1.1**



The TGal bias plot shows units of measure on the y-axis ranging from 30 µg/dL serum to -30 mg/dL blood. The bias for this plot is -1.1. The data on this plot shows variable bias across methods.

**Figure 5. Reproducibility of Results:  
Bias Plot of Thyroid-Stimulating Hormone (TSH) Values by Method  
Quarter 1 , Specimen 20201001005  
Expected Value (EV) = 85.4 µIU/mL serum**

**Quarter 1**


Specimen: 20201001005

**Enriched: 85.0**

**CDC Assayed: 79.4**

**Participant Mean: 74.2**

**Participant Bias: -11.2**



The TSH bias plot shows units of measure on the y-axis ranging from 100 µg/dL serum to -100 mg/dL blood. The bias for this plot is -11.2. The data on this plot show an even bias scatter across methods.

**Figure 6. Reproducibility of Results:  
Bias Plot of Immunoreactive Trypsinogen (IRT) Values by Method  
Quarter 1, Specimen 20201008002  
Assayed Value (AV) = 168.2 ng/mL blood**

**Quarter 1**


Specimen: 20201008002

**Enriched: 250.0**

**CDC Assayed: 168.2**

**Participant Mean: 102.9**

**Participant Bias: -65.3**



The IRT bias plot shows units of measure on the y-axis ranging from 200 µg/dL serum to -200 mg/dL blood. The bias for this plot is -65.3. This negative bias was attributed to methods with a higher participation rate.

**Figure 7. Reproducibility of Results:  
Bias Plot of Arginine (Arg) Values by Method  
Quarter 1, Specimen 20201005005  
Expected Value (EV) = 11.4 µmol/L blood**

**Quarter 1**


Specimen: 20201005005

**Enriched: 0.0**

**CDC Assayed: 11.9**

**Participant Mean: 10.2**

**Participant Bias: -1.2**



The Arg bias plot shows units of measure on the y-axis ranging from 30 µmol/L blood to -30 µmol/L blood. The bias for this plot is -1.2. This plot shows a tight scatter around the expected value across methods.

**Figure 8. Reproducibility of Results:  
Bias Plot of Citrulline (Cit) Values by Method  
Quarter 4, Specimen 20204005005  
Expected Value (EV) = 179.6  $\mu\text{mol/L}$  blood**

**Quarter 4**


Specimen: 20201005005

**Enriched: 150.0**

**CDC Assayed: 166.2**

**Participant Mean: 150.7**

**Participant Bias: -28.9**



The Cit bias plot shows units of measure on the y-axis ranging from 175  $\mu\text{mol/L}$  blood to -175  $\mu\text{mol/L}$  blood. The bias for this plot is -28.9. This plot shows a moderately negative bias across methods.

**Figure 9. Reproducibility of Results:  
Bias Plot of Leucine (Leu) Values by Method  
Quarter 1, Specimen 20201005003  
Expected Value (EV) = 620.2  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201005003

**Enriched: 470.0**

**CDC Assayed: 563.9**

**Participant Mean: 547.6**

**Participant Bias: -72.6**



The Leu bias plot shows units of measure on the y-axis ranging from 450.0  $\mu\text{mol/L}$  blood to -450.0  $\mu\text{mol/L}$  blood. The bias for this plot is -72.6. This plot shows a moderately negative bias across methods.

**Figure 10. Reproducibility of Results:  
Bias Plot of Methionine (Met) Values by Method  
Quarter 1, Specimen 20201005004  
Expected Value (EV) = 169.6  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201005004

**Enriched: 150.0**

**CDC Assayed: 139.7**

**Participant Mean: 137.3**

**Participant Bias: -32.3**



The Met bias plot shows units of measure on the y-axis ranging from 180.0  $\mu\text{mol/L}$  blood to -180.0  $\mu\text{mol/L}$  blood. The bias for this plot is -32.3. This plot shows a moderately negative bias across methods.

**Figure 11. Reproducibility of Results:  
Bias Plot of Phenylalanine (Phe) Values by Method  
Quarter 1, Specimen 20201005001  
Expected Value (EV) = 308.0  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201005001

**Enriched: 250.0**

**CDC Assayed: 298.0**

**Participant Mean: 294.5**

**Participant Bias: 13.5**



The Phe bias plot shows units of measure on the y-axis ranging from 200.0  $\mu\text{mol/L}$  blood to -200.0  $\mu\text{mol/L}$  blood. The bias for this plot is 13.5. This plot shows an even scatter across across the expected value for most methods.

**Figure 12. Reproducibility of Results:  
Bias Plot of Succinylacetone (SUAC) Values by Method  
Quarter 1, Specimen 20201005005  
Expected Value (EV) = 35.3  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201005005

**Enriched: 35.0**

**CDC Assayed: 21.0**

**Participant Mean: 15.0**

**Participant Bias: -20.3**



The SUAC bias plot shows units of measure on the y-axis ranging from 60.0  $\mu\text{mol/L}$  blood to -60.0  $\mu\text{mol/L}$  blood. The bias for this plot is -20.3. This plot shows a strongly negative bias across methods, which is historical for this analyte.

**Figure 13. Reproducibility of Results:  
Bias Plot of Tyrosine (Tyr) Values by Method  
Quarter 1, Specimen 20201005005  
Expected Value (EV) = 784.0  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201005005

**Enriched: 700.0**

**CDC Assayed: 723.9**

**Participant Mean: 705.6**

**Participant Bias: -78.4**



The Tyr bias plot shows units of measure on the y-axis ranging from 500.0  $\mu\text{mol/L}$  blood to -500.0  $\mu\text{mol/L}$  blood. The bias for this plot is -78.4. This plot shows a slightly negative bias across methods.

**Figure 14. Reproducibility of Results:  
Bias Plot of Valine (Val) Values by Method  
Quarter 1, Specimen 20201005003  
Expected Value (EV) = 599.4  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201005003

**Enriched: 450.0**

**CDC Assayed: 493.7**

**Participant Mean: 503.8**

**Participant Bias: -95.6**



The Val bias plot shows units of measure on the y-axis ranging from 500.0  $\mu\text{mol/L}$  blood to -500.0  $\mu\text{mol/L}$  blood. The bias for this plot is -95.6. This plot shows a moderately negative bias across methods.

**Figure 15. Reproducibility of Results:  
Bias Plot of Low Free Carnitine (C0(L)) Values by Method  
Quarter 1, Specimen 20201006001  
Expected Value (EV) = 7.13  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201006001

**Enriched: 0.00**

**CDC Assayed: 7.30**


**Participant Mean: 5.63**

**Participant Bias: -1.50**




The C0(L) bias plot shows units of measure on the y-axis ranging from 20.00  $\mu\text{mol/L}$  blood to -20.00  $\mu\text{mol/L}$  blood. The bias for this plot is -1.50. This plot shows a slightly negative bias across methods.

**Figure 16. Reproducibility of Results:  
Bias Plot of Low Acetylcarnitine (C2(L)) Values by Method  
Quarter 1, Specimen 20201006001  
Expected Value (EV) = 6.84  $\mu\text{mol/L}$  blood**



The C2(L) bias plot shows units of measure on the y-axis ranging from 20.00  $\mu\text{mol/L}$  blood to -20.00  $\mu\text{mol/L}$  blood. The bias for this plot is -0.14. This plot shows a tight scatter across methods.

**Figure 17. Reproducibility of Results:  
Bias Plot of Propionylcarnitine (C3) Values by Method  
Quarter 1, Specimen 20201006003  
Expected Value (EV) = 10.89  $\mu\text{mol/L}$  blood**



The C3 bias plot shows units of measure on the y-axis ranging from 8.00  $\mu\text{mol/L}$  blood to -8.00  $\mu\text{mol/L}$  blood. The bias for this plot is -0.08. This plot shows an even scatter across methods.

## Quarter 1

Specimen: 20201006001

Enriched: 0.00

CDC Assayed: 6.83

Participant Mean: 6.70

Participant Bias: -0.14

## Quarter 1

Specimen: 20201006003

Enriched: 10.00

CDC Assayed: 11.00

Participant Mean: 10.81

Participant Bias: -0.08

**Figure 18. Reproducibility of Results:  
Bias Plot of Malonylcarnitine (C3DC) Values by Method  
Quarter 1, Specimen 20201006004  
Expected Value (EV) = 11.02  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201006004

**Enriched: 11.00**

**CDC Assayed: 8.60**

**Participant Mean: 5.62**

**Participant Bias: -5.4**



The C3DC bias plot shows units of measure on the y-axis ranging from 30.00  $\mu\text{mol/L}$  blood to -30.00  $\mu\text{mol/L}$  blood. The bias for this plot is -5.4. This plot shows a slightly negative bias across methods.

**Figure 19. Reproducibility of Results:  
Bias Plot of C3DC+C4OH Non-derivatized Values by Method  
Quarter 1, Specimen 20201006004  
Expected Value (EV) = 11.03  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201006004

**Enriched: 11.00**

**CDC Assayed: 0.78**


**Participant Mean: 1.74**

**Participant Bias: -9.29**




The C3DC+C4OH bias plot shows units of measure on the y-axis ranging from 14.00  $\mu\text{mol/L}$  blood to -14.00  $\mu\text{mol/L}$  blood. The bias for this plot is -9.29. This plot shows a strongly negative bias across methods, as historically observed.

**Figure 20. Reproducibility of Results:  
Bias Plot of Butyrylcarnitine (C4) Values by Method  
Quarter 4, Specimen 20204006005  
Expected Value (EV) = 2.63  $\mu\text{mol/L}$  blood**



The C4 bias plot shows units of measure on the y-axis ranging from 3.00  $\mu\text{mol/L}$  blood to -3.00  $\mu\text{mol/L}$  blood. The bias for this plot is -0.48. This plot shows a moderately negative bias across methods.

**Figure 21. Reproducibility of Results:  
Bias Plot of Hydroxybutyrylcarnitine (C4OH) Values by Method  
Quarter 4, Specimen 20204006005  
Expected Value (EV) = 4.06  $\mu\text{mol/L}$  blood**



The C4OH bias plot shows units of measure on the y-axis ranging from 5.00  $\mu\text{mol/L}$  blood to -5.00  $\mu\text{mol/L}$  blood. The bias for this plot is -1.45. This plot shows a moderately negative bias across methods.

## Quarter 1

Specimen: 20204006005

**Enriched: 2.50**

**CDC Assayed: 2.18**

**Participant Mean: 2.15**

**Participant Bias: -0.48**

## Quarter 4

Specimen: 20204006005

**Enriched: 4.00**

**CDC Assayed: 2.86**

**Participant Mean: 2.61**

**Participant Bias: -1.45**

**Figure 22. Reproducibility of Results:  
Bias Plot of Isovalerylcarnitine (C5) Values by Method  
Quarter 4, Specimen 20204006005  
Expected Value (EV) = 2.59  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20204006005

**Enriched: 2.50**

**CDC Assayed: 2.34**

**Participant Mean: 2.15**

**Participant Bias: -0.44**



The C5 bias plot shows units of measure on the y-axis ranging from 3.00  $\mu\text{mol/L}$  blood to -3.00  $\mu\text{mol/L}$  blood. The bias for this plot is -0.44. This plot shows a moderately negative bias across methods.

**Figure 23. Reproducibility of Results:  
Bias Plot of Tiglylcarnitine (C5:1) Values by Method  
Quarter 4, Specimen 20204006004  
Expected Value (EV) = 1.02  $\mu\text{mol/L}$  blood**

**Quarter 4**


Specimen: 20204006004

**Enriched: 1.00**

**CDC Assayed: 0.74**

**Participant Mean: 0.67**

**Participant Bias: -0.35**



The C5:1 bias plot shows units of measure on the y-axis ranging from 2.00  $\mu\text{mol/L}$  blood to -2.00  $\mu\text{mol/L}$  blood. The bias for this plot is -0.35. This plot shows a moderately negative bias across methods.

**Figure 24. Reproducibility of Results:  
Bias Plot of Glutarylcarnitine (C5DC) Values by Method  
Quarter 4, Specimen 20204006003  
Expected Value (EV) = 2.05  $\mu\text{mol/L}$  blood**

**Quarter 4**


Specimen: 20204006003

**Enriched: 2.00**

**CDC Assayed: 2.05**

**Participant Mean: 1.86**

**Participant Bias: -0.19**



The C5DC bias plot shows units of measure on the y-axis ranging from 5.00  $\mu\text{mol/L}$  blood to -5.00  $\mu\text{mol/L}$  blood. The bias for this plot is -0.19. This plot shows even scatter for most methods, two methods show a positive bias and one shows a negative bias

**Figure 25. Reproducibility of Results:  
Bias Plot of Hydroxyisovalerylcarnitine (C5OH) Values by Method  
Quarter 1, Specimen 20201006005  
Expected Value (EV) = 2.56  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201006005

**Enriched: 2.00**

**CDC Assayed: 2.35**

**Participant Mean: 2.02**

**Participant Bias: -0.54**




The C5OH bias plot shows units of measure on the y-axis ranging from 4.00  $\mu\text{mol/L}$  blood to -4.00  $\mu\text{mol/L}$  blood. The bias for this plot is -0.54. This plot shows a moderately negative bias across methods.

**Figure 26. Reproducibility of Results:  
Bias Plot of Hexanoylcarnitine (C6) Values by Method  
Quarter 1, Specimen 20201006002  
Expected Value (EV) = 1.37  $\mu\text{mol/L}$  blood**

**Quarter 1**

Specimen: 20201006002

**Enriched: 1.30**  
**CDC Assayed: 1.37**  
**Participant Mean: 1.32**  
**Participant Bias: -0.05**




The C6 bias plot shows units of measure on the y-axis ranging from 1.50  $\mu\text{mol/L}$  blood to -1.50  $\mu\text{mol/L}$  blood. The bias for this plot is -0.05. This plot shows an even scatter across all methods

**Figure 27. Reproducibility of Results:  
Bias Plot of Octanoylcarnitine (C8) Values by Method  
Quarter 1, Specimen 20201006002  
Expected Value (EV) = 2.01  $\mu\text{mol/L}$  blood**

**Quarter 1**

Specimen: 20201006002

**Enriched: 1.80**  
**CDC Assayed: 2.15**  
**Participant Mean: 1.97**  
**Participant Bias: -0.04**



The C8 bias plot shows units of measure on the y-axis ranging from 3.00  $\mu\text{mol/L}$  blood to -3.00  $\mu\text{mol/L}$  blood. The bias for this plot is -0.04. This plot shows an even scatter across all methods.

**Figure 28. Reproducibility of Results:  
Bias Plot of Decanoylcarnitine (C10) Values by Method  
Quarter 1, Specimen 20201006002  
Expected Value (EV) = 1.36  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201006002

**Enriched: 1.10**

**CDC Assayed: 1.41**

**Participant Mean: 1.21**

**Participant Bias: -0.15**



The C10 bias plot shows units of measure on the y-axis ranging from 1.50  $\mu\text{mol/L}$  blood to -1.50  $\mu\text{mol/L}$  blood. The bias for this plot is -0.15. This plot shows the derivitized MassChrom kit as having a slightly more negative bias than other reported methods

**Figure 29. Reproducibility of Results:  
Bias Plot of Decenoylcarnitine (C10:1) Values by Method  
Quarter 1, Specimen 20201006002  
Expected Value (EV) = 1.08  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201006002

**Enriched: 0.90**

**CDC Assayed: 1.14**

**Participant Mean: 0.87**

**Participant Bias: -0.21**



The C10:1 bias plot shows units of measure on the y-axis ranging from 2.00  $\mu\text{mol/L}$  blood to -2.00  $\mu\text{mol/L}$  blood. The bias for this plot is -0.21. This plot shows the derivitized MassChrom kit as having a slightly more negative bias than other reported methods.

**Figure 30. Reproducibility of Results:  
Bias Plot of Decadienoylcarnitine (C10:2) Values by Method  
Quarter 4, Specimen 20204006005  
Expected Value (EV) = 0.51  $\mu\text{mol/L}$  blood**

**Quarter 4**


Specimen: 20204006005

**Enriched: 0.50**

**CDC Assayed: 0.34**

**Participant Mean: 0.25**

**Participant Bias: -0.26**



The C10:2 bias plot shows units of measure on the y-axis ranging from 0.80  $\mu\text{mol/L}$  blood to -0.80  $\mu\text{mol/L}$  blood. The bias for this plot is -0.26. This plot shows a moderately negative bias across methods.

**Figure 31. Reproducibility of Results:  
Bias Plot of Myristoylcarnitine (C14) Values by Method  
Quarter 4, Specimen 20204006001  
Expected Value (EV) = 1.71  $\mu\text{mol/L}$  blood**

**Quarter 4**


Specimen: 20204006001

**Enriched: 1.60**

**CDC Assayed: 1.71**

**Participant Mean: 1.28**

**Participant Bias: -0.43**



The C14 bias plot shows units of measure on the y-axis ranging from 2.00  $\mu\text{mol/L}$  blood to -2.00  $\mu\text{mol/L}$  blood. The bias for this plot is -0.43. This plot shows a moderately negative bias across methods.

**Figure 32. Reproducibility of Results:  
Bias Plot of Tetradecenoylcarnitine (C14:1) Values by Method  
Quarter 4, Specimen 20204006001  
Expected Value (EV) = 1.88  $\mu\text{mol/L}$  blood**

**Quarter 4**


Specimen: 20204006001

**Enriched: 1.80**

**CDC Assayed: 1.42**

**Participant Mean: 1.18**

**Participant Bias: -0.70**



The C14:1 bias plot shows units of measure on the y-axis ranging from 2.50  $\mu\text{mol/L}$  blood to -2.50  $\mu\text{mol/L}$  blood. The bias for this plot is -0.70. This plot shows a moderately negative bias across methods

**Figure 33. Reproducibility of Results:  
Bias Plot of Palmitoylcarnitine (C16) Values by Method  
Quarter 4, Specimen 20204006002  
Expected Value (EV) = 20.58  $\mu\text{mol/L}$  blood**

**Quarter 4**


Specimen: 20204006002

**Enriched: 20.00**

**CDC Assayed: 17.47**

**Participant Mean: 16.05**

**Participant Bias: -4.53**



The C16 bias plot shows units of measure on the y-axis ranging from 22.00  $\mu\text{mol/L}$  blood to -22.00  $\mu\text{mol/L}$  blood. The bias for this plot is -4.53. This plot shows a moderately negative bias across methods.

**Figure 34. Reproducibility of Results:  
Bias Plot of Hydroxypalmitoylcarnitine (C16OH) Values by Method  
Quarter 4, Specimen 20204006002  
Expected Value (EV) = 0.02  $\mu\text{mol/L}$  blood**

**Quarter 4**


Specimen: 20204006002

**Enriched: 0.0**

**CDC Assayed: 0.06**

**Participant Mean: 0.03**

**Participant Bias: 0.01**



The C16OH bias plot shows units of measure on the y-axis ranging from 0.20  $\mu\text{mol/L}$  blood to -0.20  $\mu\text{mol/L}$  blood. The bias for this plot is 0.01. This plot shows a tight scatter among all methods around the bias.

**Figure 35. Reproducibility of Results:  
Bias Plot of Stearoylcarnitine (C18) Values by Method  
Quarter 4, Specimen 20204006002  
Expected Value (EV) = 5.47  $\mu\text{mol/L}$  blood**

**Quarter 4**


Specimen: 20204006002

**Enriched: 5.00**

**CDC Assayed: 5.21**

**Participant Mean: 4.73**

**Participant Bias: -0.74**



The C18 bias plot shows units of measure on the y-axis ranging from 3.50  $\mu\text{mol/L}$  blood to -3.50  $\mu\text{mol/L}$  blood. The bias for this plot is -0.74. This plot shows a moderately negative bias across methods.

**Figure 36. Reproducibility of Results:  
Bias Plot of Oleoylcarnitine (C18:1) Values by Method  
Quarter 4, Specimen 20204006002  
Expected Value (EV) = 11.52  $\mu\text{mol/L}$  blood**

**Quarter 4**


Specimen: 20204006002

**Enriched: 10.00**

**CDC Assayed: 9.32**

**Participant Mean: 7.87**

**Participant Bias: -3.65**



The C18:1 bias plot shows units of measure on the y-axis ranging from 12.00  $\mu\text{mol/L}$  blood to -12.00  $\mu\text{mol/L}$  blood. The bias for this plot is -3.65. This plot shows a moderately negative bias across methods.

**Figure 37. Reproducibility of Results:  
Bias Plot of Hydroxystearoylcarnitine (C18OH) Values by Method  
Quarter 1, Specimen 20201006002  
Expected Value (EV) = 0.01  $\mu\text{mol/L}$  blood**

**Quarter 1**


Specimen: 20201006002

**Enriched: 0.00**

**CDC Assayed: 0.01**

**Participant Mean: 0.01**

**Participant Bias: 0.00**



The C18OH bias plot shows units of measure on the y-axis ranging from 0.12  $\mu\text{mol/L}$  blood to -0.12  $\mu\text{mol/L}$  blood. The bias for this plot is 0.00. The derivatized MS/MS non-kit method shows a slightly higher bias than other methods.

## Appendix for Accessibility Descriptions

**Figures 2–37, Bias Plots:** Bias plots, which compare two measurements of the same variable, have been created to show a wide range of PT challenge specimens. The bias, which is calculated by subtracting the participant mean value from the CDC expected value (EV), is represented by the broken line. The EV is the sum of the endogenous plus the enrichment values. The solid line represents perfect agreement with the EV or zero bias. When comparing data scatter among figures, the scale (y-axis) might differ. We included the 95% confidence interval for the mean participant bias. A tight scatter within this interval indicates good performance for a method or a group of methods. To illustrate any method-related differences in analyte recoveries, we grouped the PT quantitative results by kit or method. Because some of the pools in a routine PT survey represent a unique donor specimen, differences in endogenous materials in the donor specimens might influence method-related differences. We showed representative bias plots for all those analytes distributed in PT challenges that required a quantitative measurement to determine the presumptive clinical assessments.

## References

[1] Newborn Screening: Towards a Uniform Screening Panel and System." *Genetic Medicine* 2006;8(5) Suppl: S12–S252, as authored by the American College of Medical Genetics and commissioned by the Health Resources and Services Administration.

[2] Clinical and Laboratory Standards Institute. Blood collection on filter paper for newborn screening programs: Approved Standard—Sixth Edition. CLSI Document NBS01-A6. Wayne, PA: Clinical and Laboratory Standards Institute; 2014.

# Acknowledgments

This NEWBORN SCREENING QUALITY ASSURANCE PROGRAM report is an internal publication distributed to program participants and selected program colleagues. The laboratory quality assurance program is a project cosponsored by the Centers for Disease Control and Prevention (CDC) and the Association of Public Health Laboratories.

## CENTERS FOR DISEASE CONTROL AND PREVENTION (CDC) ATLANTA, GA 30341

### Director

Robert R. Redfield, M.D.

### Director

National Center for Environmental Health  
Patrick Breyse, Ph.D.

### Director

Division of Laboratory Sciences  
James L. Pirkle, M.D., Ph.D.

### Chief

Newborn Screening and Molecular Biology Branch  
Carla Cuthbert, Ph.D.

---

### Contributors

Nicole Baird, PhD  
John Bernstein, MS  
Christofer Brown, BS  
Laura Centeno Badilo Hildreth, BS  
Suzanne Cordovado, PhD  
Paul Dantonio, MS  
Katherine Duneman, MS  
Elise Gowen, MS  
Christopher Greene, PhD  
Laura Hancock, MS  
Christopher Haynes, PhD  
Cynthia Hinton, PhD  
Samantha Isenberg, PhD

Matt Kilgore, PhD  
Deborah Koontz, PhD  
Francis Lee, PhD  
LiXia Li, PhD  
Tim Lim, PhD  
Edgardo Lobo, MS  
Allison McCabe, BS  
Elizabeth McCown, BS  
Joanne Mei, PhD  
Kristina Mercer, PhD  
Stanimila Nikolova, PhD  
Kaila Pearson, MS  
Gyliann Pena, MS

Kostas Petritis, PhD  
C. Austin Pickens, PhD  
Maryam Salehi, PhD  
Blanche Temate, PhD  
Robert Vogt, PhD  
Irene Williams, MS  
Golriz Yazdanpanah, MS  
Sherri Zobel, BS

### Production

Vinay Anumula, MS  
Kizzy Stewart  
Dion Cannon

---

## ASSOCIATION OF PUBLIC HEALTH LABORATORIES SILVER SPRING, MD 20910

### President

Grace E. Kubin, PhD

### Chairman, Newborn Screening and Genetics in Public Health Committee

Michele Caggana, Sc.D., FACMG

### Chairman, Newborn Screening Quality Assurance Quality Control Subcommittee

Patricia R. Hunt, B.A. and Joseph Orsini, Ph.D.

### Chairman, Newborn Screening Molecular Subcommittee

Rachel Lee, Ph.D.

---

### INQUIRIES TO:

Sherri Zobel, Editor  
Centers for Disease Control and Prevention (CDC),  
Newborn Screening Quality Assurance Program—Mailstop S110-3  
4770 Buford Highway, N.E., Atlanta, GA 30341-3724  
E-mail: [NSQAPDMT@cdc.gov](mailto:NSQAPDMT@cdc.gov)



**For more information please contact  
Centers for Disease Control and Prevention**

1600 Clifton Road NE, Atlanta, GA 33029-4027

Telephone: 1-800-CDC-INFO (232-4636)

TTY: 1-888-232-6348

E-mail: [NSQAPDMT@cdc.gov](mailto:NSQAPDMT@cdc.gov)

Web: <https://www.cdc.gov/labstandards/nsqap.html>

Publication date: May 2021